

AMENDMENTS TO THE CLAIMS:

Claim 5 has been withdrawn. Claims 1 and 13-15 have been amended, as show below.

This listing of claims will replace all prior versions and listings of claims in the
Application:

Claim 1 (currently amended): A blister pack for use with inhalation therapy inhalers equipped with a vibratory de-aggregator, comprising a first element comprising an elongate flexible coiled tape having spaced areas for interfacing and coupling with ~~the~~ a inhaler vibratory de-aggregator, and a frangible second element overlying the first element and defining a plurality of spaced top crowned areas located over said first element spaced areas, and containing powder or liquid material, wherein the contents of the blister are forcibly ejected from the blister through puncture holes formed in the top crown areas when the spaced areas are interfaced coupled with ~~the~~ a vibratory de-aggregator.

Claim 2 (cancelled)

Claim 3 (original): A blister pack according to claim 1, wherein said top crowned areas are shaped as inverted cones.

Claim 4 (withdrawn)

Claim 5 (withdrawn): A blister pack according to claim 1, wherein the first element includes a depression opposite the top crowned areas.

Claims 6-7 (withdrawn)

Claim 8. (original): A blister pack according to claim 1, wherein said material comprises a medication.

HAYES SOLOWAY P.C.
130 W. CUSHING ST.
TUCSON, AZ 85701
TEL. 520.882.7623
FAX. 520.882.7643

175 CANAL STREET
MANCHESTER, NH 03101
TEL. 603.668.1400
FAX. 603.668.8567

Claim 9 (original): A blister pack according to claim 1, wherein said material comprises a vitamin.

Claim 10 (original): A blister pack according to claim 1, wherein said material comprises a hormone.

Claim 11 (original): A blister pack according to claim 1, wherein said material comprises a steroid.

Claim 12 (original): A blister pack according to claim 1, wherein said material comprises a bioactive material.

Claim 13 (currently amended): A blister pack according to claim 1, wherein ~~the~~ a size and number of puncture holes together with volume formed by the blister pack are optimized for de-aggregation and aerosolization of material in the blister pack.

Claim 14 (currently amended): A blister pack according to claim 1, wherein ~~the~~ a height and shape of the blister pack is optimized for de-aggregation and aerosolization of material in the blister pack.

Claim 15 (currently amended): A blister pack according to claim 1, wherein ~~the~~ an interface to ~~the~~ a vibrator is optimized for optimum coupling of the energy into the blister pack for de-aggregation and aerosolization of material in the blister pack.

HAYES SOLOWAY P.C.
130 W. CUSHING ST.
TUCSON, AZ 85701
TEL. 520.882.7623
FAX. 520.882.7643

175 CANAL STREET
MANCHESTER, NH 03101
TEL. 603.668.1400
FAX. 603.668.8567